

## Clinical Biochemistry Lab

Requirement for Academic session July 2009 to June 2010

### NON CONSUMABLE

#### 1- HOT AIR OVEN-----01

##### Specifications

- Oven with door mounted components
- Temperature range- 20 to 200°C
- Temperature control accuracy- $\pm 0.1$  C
- Temperature display- alphanumeric LCD
- Resolution-0.1°C
- Timer- 1 min. to 999 hr 59 min.
- Control type- microprocessor
- Construction-Interior-stainless steel  
Exterior-powder coated
- Capacity- 1.4 cu feet
- Temperature uniformity-- $\pm 0.7^\circ\text{C}$  at 150°C
- Dimensions Chamber-13<sup>3/8"</sup>x14<sup>5/8"</sup>x15<sup>1/2"</sup> inches
- Power- 230 volt AC, 50/60Hz

#### 2- INCUBATOR-----01

##### Specifications

- Acrylic single walled see through type
- Temperature control range- (ambient+5) to 60°C
- Temperature control accuracy- $\pm 0.2^\circ\text{C}$